

ABX Minotrol CRP



LOT

MC 433

CONTROL



(Exp.) 2025-01-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0		2.1		± 0.4	8.0	7.8		8.1		± 1.0	20.1	20.4		21.2		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.37	2.38		2.34		± 0.15	4.68	4.69		4.67		± 0.18	5.56	5.60		5.64		± 0.25
HB	HGB	g/dl	6.2	6.2		6.0		± 0.4	13.6	13.6		13.8		± 0.5	17.9	18.1		18.4		± 0.7
		g/l	62	62		60		± 4	136	136		138		± 5	179	181		184		± 7
		mmol/l	3.85	3.85		3.73		± 0.25	8.45	8.45		8.57		± 0.31	11.12	11.24		11.43		± 0.44
HT	HCT	%	16.6	16.7		16.6		± 2.0	37.4	37.1		37.4		± 2.5	50.0	49.8		49.9		± 3.0
		l/l	0.166	0.167		0.166		± 0.020	0.374	0.371		0.374		± 0.025	0.500	0.498		0.499		± 0.030
VGM	MCV	µm ³ ; fl	70	70		71.0		± 4	80	79		80.0		± 4	90	89		88.5		± 4
TGMH	MCH	pg	26.2	26.1		25.6		± 2.4	29.1	29.0		29.6		± 2.4	32.2	32.3		32.6		± 2.8
		fmol	1.62	1.62		1.59		± 0.15	1.80	1.80		1.84		± 0.15	2.00	2.01		2.03		± 0.18
CCMH	MCHC	g/dl	37.4	37.2		36.1		± 3.0	36.3	36.7		36.9		± 3.0	35.8	36.3		36.9		± 3.0
		g/l	374	372		361		± 30	363	367		369		± 30	358	363		369		± 30
		mmol/l	23.21	23.11		22.43		± 1.86	22.56	22.79		22.94		± 1.86	22.21	22.55		22.89		± 1.86
IDR	RDW	%	13.5	13.5		14.5		± 3.0	14.5	14.0		12.5		± 3.0	14.0	13.5		11.0		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	67	76		63		± 20	237	249		243		± 40	480	511		507		± 60
VPM	MPV	µm ³ ; fl	7.8	8.4		9.3		± 2.0	7.9	8.3		9.0		± 2.0	7.9	8.1		8.8		± 2.0
LYMPHO		#	1.3	1.3		N/A		± 0.2	2.8	2.8		N/A		± 0.5	3.3	3.3		N/A		± 1.1
		%	67.0	66.0		N/A		± 8.0	34.5	35.5		N/A		± 6.0	16.5	16.0		N/A		± 5.0
MONO		#	0.2	0.2		N/A		± 0.2	0.5	0.5		N/A		± 0.5	1.5	1.4		N/A		± 1.3
		%	7.5	8.0		N/A		± 6.8	6.5	6.0		N/A		± 5.4	7.5	7.0		N/A		± 6.3
GRANULO		#	0.5	0.5		N/A		± 0.4	4.7	4.6		N/A		± 0.8	15.3	15.7		N/A		± 2.0
		%	25.5	26.0		N/A		± 10.0	59.0	58.5		N/A		± 10.0	76.0	77.0		N/A		± 10.0
CRP		mg/dl	0.47	0.55		0.56		± 0.40	2.40	2.92		2.81		± 0.80	4.56	5.29		5.42		± 1.50
		mg/l	4.7	5.5		5.6		± 4.0	24.0	29.2		28.1		± 8.0	45.6	52.9		54.2		± 15.0

Ref: TEMP-1070 Rev.43 BACK/VERSO 1300092170