

ABX Minotrol CRP



LOT

MC 436

CONTROL



(Exp.) 2025-10-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																						
				CONTROL				1	TOLERANCES TOLERANCE		CONTROL				2		TOLERANCES TOLERANCE		CONTROL				3		TOLERANCES TOLERANCE	
				MICROS			PENTRA				MICROS			PENTRA		MICROS			PENTRA							
				EMI CRP	LC-767G		MS CRP				EMI CRP	LC-767G		MS CRP		EMI CRP			LC-767G		MS CRP					
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.0		2.1		± 0.4	8.0	7.8		8.0		± 1.0	20.4	20.4		21.0		± 2.2						
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.22	2.22		2.13		± 0.15	4.54	4.50		4.47		± 0.18	5.47	5.47		5.50		± 0.25						
HB	HGB	g/dl	6.1	6.1		5.9		± 0.4	13.3	13.3		13.4		± 0.5	17.6	17.8		18.0		± 0.7						
		g/l	61	61		59		± 4	133	133		134		± 5	176	178		180		± 7						
		mmol/l	3.79	3.79		3.66		± 0.25	8.26	8.26		8.32		± 0.31	10.93	11.05		11.18		± 0.44						
HT	HCT	%	16.7	16.4		16.2		± 2.0	36.3	36.0		36.2		± 2.5	48.7	48.7		48.4		± 3.0						
		l/l	0.167	0.164		0.162		± 0.020	0.363	0.360		0.362		± 0.025	0.487	0.487		0.484		± 0.030						
VGM	MCV	µm ³ ; fl	75	74		76.0		± 4	80	80		81.0		± 4	89	89		88.0		± 4						
TGMH	MCH	pg	27.5	27.5		27.7		± 2.4	29.3	29.6		30.0		± 2.4	32.2	32.5		32.7		± 2.8						
		fmol	1.71	1.71		1.72		± 0.15	1.82	1.84		1.86		± 0.15	2.00	2.02		2.03		± 0.18						
CCMH	MCHC	g/dl	36.6	37.1		36.4		± 3.0	36.6	36.9		37.0		± 3.0	36.2	36.6		37.2		± 3.0						
		g/l	366	371		364		± 30	366	369		370		± 30	362	366		372		± 30						
		mmol/l	22.75	23.06		22.63		± 1.86	22.74	22.94		22.98		± 1.86	22.45	22.71		23.10		± 1.86						
IDR	RDW	%	14.5	14.0		14.5		± 3.0	14.0	13.5		12.5		± 3.0	14.0	13.5		11.5		± 3.0						
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	75	74		70		± 20	259	252		264		± 40	483	486		513		± 60						
VPM	MPV	µm ³ ; fl	7.2	7.9		8.2		± 2.0	7.2	7.3		7.5		± 2.0	7.3	7.4		7.4		± 2.0						
LYMPHO	#		1.3	1.3		N/A		± 0.2	2.7	2.6		N/A		± 0.5	3.4	3.2		N/A		± 1.1						
	%		63.5	63.0		N/A		± 8.0	33.5	33.0		N/A		± 6.0	16.5	15.5		N/A		± 5.0						
MONO	#		0.2	0.2		N/A		± 0.2	0.6	0.5		N/A		± 0.5	1.2	1.1		N/A		± 1.0						
	%		8.5	8.0		N/A		± 7.2	7.0	7.0		N/A		± 6.3	6.0	5.5		N/A		± 5.0						
GRANULO	#		0.6	0.6		N/A		± 0.2	4.8	4.7		N/A		± 0.8	15.8	16.1		N/A		± 2.0						
	%		28.0	29.0		N/A		± 9.9	59.5	60.0		N/A		± 9.9	77.5	79.0		N/A		± 9.9						
CRP	mg/dl		0.41	0.49		0.45		± 0.40	2.27	2.48		2.48		± 0.80	4.37	5.14		4.91		± 1.50						
	mg/l		4.1	4.9		4.5		± 4.0	22.7	24.8		24.8		± 8.0	43.7	51.4		49.1		± 15.0						

1300092170
BACK/VERSO
Ref: TEMP-1070 Rev.43