

# ichroma MAU Control

## Afias

LOT MACPF47 Verfall: 2020.06

### MAU

<b>Max. Toleranzbreite QUALAB</b>	24%
<b>Toleranzbreite Hersteller</b>	24%
<b>Zielwert (mg/l)</b>	62
<b>Standardabweichung (mg/l)</b>	4.96

<b>Einheit</b>	<b>mg/l</b>
<b>Datum</b>	<b>Messwert</b>


<b>-3s</b>	<b>-2s</b>	<b>-1s</b>	<b>Mw</b>	<b>+1s</b>	<b>+2s</b>	<b>+3s</b>
47.12	52.08	57.04	62.00	66.96	71.92	76.88