

ABX Minotrol CRP


LOT
MC 438
CONTROL

(Exp.) 2026-04-05

(YYYY - MM - DD)

Rev 2

ABX Lysebio

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio																	
			CONTROL				1	TOLERANCES TOLERANCE	CONTROL				2	TOLERANCES TOLERANCE	CONTROL				3	TOLERANCES TOLERANCE
			MICROS			PENTRA			MICROS			PENTRA			MICROS			PENTRA		
			EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G		MS CRP		
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.2	2.1		2.2		± 0.4	8.0	7.8		8.1		± 1.0	20.6	20.6		21.5		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.26	2.22		2.19		± 0.15	4.54	4.46		4.53		± 0.18	5.56	5.55		5.61		± 0.25
HB	HGB	g/dl	6.0	5.9		5.7		± 0.4	13.4	13.3		13.6		± 0.5	17.4	17.6		18.2		± 0.7
		g/l	60	59		57		± 4	134	133		136		± 5	174	176		182		± 7
		mmol/l	3.73	3.66		3.54		± 0.25	8.32	8.26		8.45		± 0.31	10.81	10.93		11.30		± 0.44
HT	HCT	%	16.3	16.0		15.8		± 2.0	37.2	36.6		37.1		± 2.5	48.9	48.8		49.4		± 3.0
		l/l	0.163	0.160		0.158		± 0.020	0.372	0.366		0.371		± 0.025	0.489	0.488		0.494		± 0.030
VGM	MCV	µm ³ ; fl	72	72		72.0		± 4	82	82		82.0		± 4	88	88		88.0		± 4
TGMH	MCH	pg	26.5	26.6		26.0		± 2.4	29.5	29.8		30.0		± 2.4	31.3	31.7		32.4		± 2.8
		fmol	1.65	1.65		1.62		± 0.15	1.83	1.85		1.86		± 0.15	1.94	1.97		2.01		± 0.18
CCMH	MCHC	g/dl	36.9	36.9		36.1		± 3.0	36.0	36.4		36.6		± 3.0	35.6	36.0		36.9		± 3.0
		g/l	369	369		361		± 30	360	364		366		± 30	356	360		369		± 30
		mmol/l	22.90	22.92		22.45		± 1.86	22.35	22.58		22.74		± 1.86	22.08	22.38		22.89		± 1.86
IDR	RDW	%	15.0	15.0		15.0		± 3.0	14.0	13.5		13.0		± 3.0	14.0	13.5		12.0		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	69	71		65		± 20	273	273		267		± 40	483	491		503		± 60
VPM	MPV	µm ³ ; fl	7.8	8.4		8.9		± 2.0	8.1	8.2		8.8		± 2.0	7.7	7.9		8.4		± 2.0
LYMPHO	#	1.4	1.4		N/A		± 0.2	2.6	2.5		N/A		± 0.5	3.2	3.1		N/A		± 1.1	
	%	65.5	65.0		N/A		± 8.0	33.0	32.5		N/A		± 6.0	15.5	15.0		N/A		± 5.0	
MONO	#	0.2	0.2		N/A		± 0.2	0.6	0.5		N/A		± 0.5	1.2	1.1		N/A		± 1.0	
	%	8.0	8.0		N/A		± 7.2	7.5	7.0		N/A		± 6.3	6.0	5.5		N/A		± 5.0	
GRANULO	#	0.6	0.6		N/A		± 0.2	4.8	4.7		N/A		± 0.8	16.2	16.4		N/A		± 2.1	
	%	26.5	27.0		N/A		± 9.9	59.5	60.5		N/A		± 9.9	78.5	79.5		N/A		± 9.9	
CRP	mg/dl	1.05	1.12		1.07		± 0.40	2.69	2.69		2.77		± 0.80	5.05	4.78		5.46		± 1.50	
	mg/l	10.5	11.2		10.7		± 4.0	26.9	26.9		27.7		± 8.0	50.5	47.8		54.6		± 15.0	

Ref: TEMP-1070 Rev.43 BACK/VERSO 1300092170