

ABX Minotrol CRP



LOT MC 439
Rev 2'

CONTROL

(Exp.) 2026-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio														
			CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE		
			MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA	TOLERANCES TOLERANCE			
			EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0	2.0	± 0.4	7.7	7.6	7.7	± 1.0	20.1	20.3	20.7	± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.50	2.50	2.41	± 0.15	4.83	4.74	4.75	± 0.18	5.74	5.73	5.75	± 0.25			
HB	HGB	g/dl	6.4	6.4	6.2	± 0.4	13.9	13.9	14.1	± 0.5	18.2	18.3	18.7	± 0.7			
		g/l	64	64	62	± 4	139	139	141	± 5	182	183	187	± 7			
		mmol/l	3.97	3.97	3.85	± 0.25	8.63	8.63	8.76	± 0.31	11.30	11.36	11.61	± 0.44			
HT	HCT	%	17.5	17.3	17.1	± 2.0	38.6	37.9	38.0	± 2.5	51.1	50.4	50.0	± 3.0			
		l/l	0.175	0.173	0.171	± 0.020	0.386	0.379	0.380	± 0.025	0.511	0.504	0.500	± 0.030			
VGM	MCV	µm ³ ; fl	70	69	71.0	± 4	80	80	80.0	± 4	89	88	87.0	± 4			
TGMH	MCH	pg	25.6	25.6	25.7	± 2.4	28.8	29.3	29.7	± 2.4	31.7	31.9	32.5	± 2.8			
		fmol	1.59	1.59	1.60	± 0.15	1.79	1.82	1.84	± 0.15	1.97	1.98	2.02	± 0.18			
CCMH	MCHC	g/dl	36.6	37.1	36.2	± 3.0	36.0	36.7	37.1	± 3.0	35.6	36.3	37.4	± 3.0			
		g/l	366	371	362	± 30	360	367	371	± 30	356	363	374	± 30			
		mmol/l	22.71	23.04	22.50	± 1.86	22.34	22.76	23.04	± 1.86	22.12	22.54	23.21	± 1.86			
IDR	RDW	%	13.5	14.0	14.5	± 3.0	14.0	13.5	12.5	± 3.0	14.0	13.5	11.5	± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	71	74	65	± 20	244	243	246	± 40	491	485	490	± 60			
VPM	MPV	µm ³ ; fl	7.5	8.4	9.5	± 2.0	7.3	7.6	8.7	± 2.0	7.8	8.1	8.8	± 2.0			
LYMPHO	#		1.2	1.2	N/A	± 0.2	2.5	2.4	N/A	± 0.5	3.1	3.0	N/A	± 1.1			
	%		62.0	61.0	N/A	± 8.0	32.5	32.0	N/A	± 6.0	15.5	15.0	N/A	± 5.0			
MONO	#		0.2	0.2	N/A	± 0.2	0.6	0.6	N/A	± 0.5	1.9	1.8	N/A	± 1.6			
	%		9.0	9.5	N/A	± 8.1	7.5	7.5	N/A	± 6.8	9.5	9.0	N/A	± 8.1			
GRANULO	#		0.6	0.6	N/A	± 0.2	4.6	4.6	N/A	± 0.8	15.1	15.4	N/A	± 2.0			
	%		29.0	29.5	N/A	± 9.9	60.0	60.5	N/A	± 9.9	75.0	76.0	N/A	± 9.9			
CRP	mg/dl		0.88	0.89	0.91	± 0.40	2.30	2.40	2.19	± 0.80	3.97	3.89	4.19	± 1.50			
	mg/l		8.8	8.9	9.1	± 4.0	23.0	24.0	21.9	± 8.0	39.7	38.9	41.9	± 15.0			

Ref: TEMP-1070 Rev.43 BACKVERSO 1300092170