

ABX Minotrol CRP



LOT

MC 440

Rev 1

CONTROL



(Exp.) **2026-10-05**

(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	1.9	1.9		1.9			± 0.4	7.6	7.5			7.8		± 1.0	19.7	19.9		
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.30	2.31		2.22		± 0.15	4.46	4.46		4.45		± 0.18	5.62	5.66		5.67		± 0.25
HB	HGB	g/dl	6.1	6.1		5.7		± 0.4	13.3	13.3		13.4		± 0.5	17.7	17.9		18.3		± 0.7
		g/l	61	61		57		± 4	133	133		134		± 5	177	179		183		± 7
		mmol/l	3.79	3.79		3.54		± 0.25	8.26	8.26		8.32		± 0.31	10.99	11.12		11.36		± 0.44
HT	HCT	%	16.6	16.4		15.8		± 2.0	36.6	36.1		36.0		± 2.5	50.0	49.8		49.3		± 3.0
		l/l	0.166	0.164		0.158		± 0.020	0.366	0.361		0.360		± 0.025	0.500	0.498		0.493		± 0.030
VGM	MCV	µm ³ ; fl	72	71		71.0		± 4	82	81		81.0		± 4	89	88		87.0		± 4
TGMH	MCH	pg	26.5	26.4		25.7		± 2.4	29.8	29.8		30.1		± 2.4	31.5	31.6		32.3		± 2.8
		fmol	1.65	1.64		1.59		± 0.15	1.85	1.85		1.87		± 0.15	1.96	1.96		2.00		± 0.18
CCMH	MCHC	g/dl	36.8	37.2		36.2		± 3.0	36.4	36.8		37.2		± 3.0	35.4	35.9		37.1		± 3.0
		g/l	368	372		362		± 30	364	368		372		± 30	354	359		371		± 30
		mmol/l	22.88	23.10		22.46		± 1.86	22.58	22.86		23.09		± 1.86	21.98	22.32		23.04		± 1.86
IDR	RDW	%	15.5	15.5		15.5		± 3.0	14.0	13.5		12.5		± 3.0	14.0	13.0		12.0		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	65	69		60		± 20	247	254		253		± 40	459	482		487		± 60
VPM	MPV	µm ³ ; fl	7.3	8.1		8.7		± 2.0	7.7	8.0		8.6		± 2.0	7.6	7.9		8.2		± 2.0
LYMPHO		#	1.2	1.1		N/A		± 0.2	2.4	2.4		N/A		± 0.5	3.1	3.0		N/A		± 1.1
		%	62.0	60.5		N/A		± 8.0	32.0	31.5		N/A		± 6.0	15.5	15.0		N/A		± 5.0
MONO		#	0.2	0.2		N/A		± 0.2	0.6	0.6		N/A		± 0.5	1.3	1.3		N/A		± 1.2
		%	10.0	11.0		N/A		± 9.0	8.5	8.5		N/A		± 7.7	6.5	6.5		N/A		± 5.9
GRANULO		#	0.5	0.5		N/A		± 0.2	4.5	4.5		N/A		± 0.8	15.4	15.6		N/A		± 2.0
		%	28.0	28.5		N/A		± 9.9	59.5	60.0		N/A		± 9.9	78.0	78.5		N/A		± 9.9
CRP		mg/dl	0.87	0.97		0.92		± 0.40	2.13	2.34		2.34		± 0.80	3.92	4.26		4.57		± 1.50
		mg/l	8.7	9.7		9.2		± 4.0	21.3	23.4		23.4		± 8.0	39.2	42.6		45.7		± 15.0

Ref: TEMP-1070 Rev.43 BACKVERSO 1300092170